

# ORDERING INFORMATION

# MODEL : 40DT

## PLEASE FILL IN THIS SECTION



Model
Company
Name
P/O No.

## FACTORY USE ONLY



Job No.	Approved by (Sales office)
Ser No.	Issued by (Sales office)
Sales	Approved by (Factory)
	Set by (Factory)
	Ser No.

**Specify the items you want to change. Default setting will be used if not specified.**

DEFAULT shows values in case of nothing specified.

### ■ ZERO & SPAN ADJUSTMENT

ITEM	AVAILABLE VALUE	FUNCTION	INPUT CODE	DEFAULT VALUE	SET VALUE	Factory Internal check
Input type	K	K thermocouple (usable range -150 to 1370 °C, -238 to 2498 °F)	T1	K	<input type="checkbox"/> K	<input type="checkbox"/> Checked
	J	J thermocouple (usable range -180 to 1000 °C, -292 to 1832 °F)			<input type="checkbox"/> J	
	T	T thermocouple (usable range -170 to 400 °C, -247 to 752 °F)			<input type="checkbox"/> T	
	R	R thermocouple (usable range 380 to 1760 °C, 716 to 3200 °F)			<input type="checkbox"/> R	
Temperature Unit	C	C: °C	T1	C	<input type="checkbox"/> C	<input type="checkbox"/> Checked
	F	F: °F			<input type="checkbox"/> F	

### ■ DISPLAY SETTING

ITEM	DISPLAY	FUNCTION	DEFAULT VALUE	SET VALUE	Factory Internal check
Moving Average	AoFF	No moving averaging	A4	<input type="checkbox"/> AoFF	<input type="checkbox"/> Checked
	A2	Moving average with 2 samples		<input type="checkbox"/> A2	
	A4	Moving average with 4 samples		<input type="checkbox"/> A4	
	A8	Moving average with 8 samples		<input type="checkbox"/> A8	
	A16	Moving average with 16 samples		<input type="checkbox"/> A16	
Brightness	C1	Brightness level 1 (dark)	C5	<input type="checkbox"/> C1	<input type="checkbox"/> Checked
	C2	Brightness level 2		<input type="checkbox"/> C2	
	C3	Brightness level 3		<input type="checkbox"/> C3	
	C4	Brightness level 4		<input type="checkbox"/> C4	
	C5	Brightness level 5 (bright)		<input type="checkbox"/> C5	
CJC Sensor	Con	CJC sensor Enabled	Con	<input type="checkbox"/> Con	<input type="checkbox"/> Checked
	Coff	CJC sensor Disabled		<input type="checkbox"/> Coff	

■ ENGINEERING UNIT STICKER LABEL

ITEM	AVAILABLE RANGE		SET VALUE	DEFAULT VALUE	Factory Internal check
<b>Engineering unit</b>	DC, AC, mV, V, kV, $\mu$ A, mA, A, kA, mW, W, kW, var, kvar, Mvar, VA, Hz, $\Omega$ , k $\Omega$ , M $\Omega$ , cm, mm, m, m/sec, mm/min, cm/min, m/min, m/h, m/s <sup>2</sup> , inch, l, l/s, l/min, l/h, m <sup>3</sup> , m <sup>3</sup> /sec, m <sup>3</sup> /min, m <sup>3</sup> /h, Nm <sup>3</sup> /h, N·m, N/m <sup>2</sup> , g, kg, kg/h, N, kN, Pa, kPa, MPa, t, t/h, °C, °F, %RH, J, kJ, MJ, rpm, sec, min, pH, %, ppm	Select one if necessary.		none	<input type="checkbox"/> Checked

\*Location of the sticker label

